

Critical Operational Device Specification (CODS)

1. **Critical Specifications:** The below are critical specifications and the *minimum must be met* for the device to be considered acceptable by the government. They are separated by “minimum” and “desired” columns.
2. **Instructions:** Indicate “yes” or “no” under both minimum and desired columns below. Additional comments may be added.

3. Critical Operational Device Specifications

#	Operational Characteristic	Minimum Required	Desired Characteristics	Meets Minimum Y/N?	Meets Desired Y/N?	Comments
1	Usage	The vendor shall provide a Beckman Coulter MicroScan autoSCAN-4 system software LabPro V5.0 System.	Minimum Required		N/A	
2	Positive Panel	The system shall support B1017-219 Positive Combo 45.	Minimum Required		N/A	
3	Negative Panel	The system shall support B1017-436 Negative Urine Combo 86	Minimum Required		N/A	
4	Power	The system shall operate on 115 V and 60/50 Hz.	The system operates on. 115 V and 220V with 60/50 Hz.			
5	Computer Type	The vendor shall provide a Beckman Coulter part # C52587 Labpro Computer W/Monitor with windows 11.	Minimum Required		N/A	

Critical Operational Device Specification (CODS)

4. Critical Logistic Specifications

#	Operational Characteristic	Minimum Required	Desired Characteristics	Meets Minimum /N?	Meets Desired Y/N?	Comments
1	Quality Control Material	The vendor shall provide the following quality control, hardware and consumables: ▪ B1018-14 RENOK REHYDRATOR ▪ B1010-56B CVR TRAYS 96 WELL US(12) ▪ B1013-40 INOCULATORS-D USA 240/PK ▪ 81026-10D PROMPT INOCULATION SYSTEM-D (60)	Minimum Required		N/A	
2	Training	The vendor shall provide Beckman Coulter customer training C35924 customer training Autoscan-4	Minimum Required		N/A	
3	Service Parts	The vendor shall provide the following service parts: ▪ 9800-2706 ASSAY PREVENTIVE MAINTENANCE KIT AS4 ▪ 5400-2925 HINGE CASE COVER AS4 ▪ 5029-0671 SCREWDRIVER PHILLIPS 6-32 X ¾	Minimum Required		N/A	

Critical Operational Device Specification (CODS)

		▪ C32992 SERVICE KIT AS4 3A FUSE AND LABEL				
--	--	--	--	--	--	--

5. ADDITIONAL PERFORMANCE PARAMETERS

These are desired Additional Performance Parameters (APPs) but are not essential.

#	Additional Performance Parameters (APPs)	Meets Additional Performance Parameters (APPs) Y/N?
1	System has Common Access Card (CAC) enabled with Active Client.	

End of CODS